



☒ **CTL Provisional DECISION SHEET (PDSH)**
☐ **CTL DECISION SHEET (DSH)**

Standard(s) (incl. year)	Subclause(s)	Tracking No.
OD-5011 ed.2.0	4	2297
Category	8.1.1 c)	
General		
Subject	Keywords	Developed by
Clarification of requirements	Calibration	WG 1
Question		
1. The definition of Calibration reads: Operation providing metrological traceability of measuring instrument by means of an unbroken chain of calibrations or comparisons with measuring standards or reference materials, linking them to relevant primary standards of the SI units of measurement. It is unusual to include the term we are trying to define in the text of definition. Is this intended? 2. Clause 8.1.1 c) is about the NCB/CBTL assessment to the non-accredited laboratories when using them. The note says "An on-site assessment for the external non-accredited laboratory is not required." Does it mean such external non-accredited lab need not be covered by IECEE on- site peer assessment, or a remote assessment from NCB/CBTL to such external lab is sufficient?		
Decision		
Ad: 1) Text is unclear. OD 5011 will be revised to ed. 2.1 Ad. 2.) Yes, the competency of the external not accredited calibration laboratory must be ensured by the NCB/CBTL .		
Explanatory notes		
Enquiry: 2297 Date: 2025-09-08		
PDSH: Year of publication: 2025 To be approved at CTL plenary: 2026		
DSH Publication date:		
Click here to enter text.		